

Type: RD-700

## CLINICAL LABORATORY REFRACTOMETER



CE



High precision digital refractometer for clinical use, especially for Urine specific gravity measurement. The instrument is CE approved.

### Specifications

Approval	CE approved
Display Digital	Yes
Specific Gravity	
Measuring Range	1.000-1.090
Minimum Scale	±0.001
Accuracy	±0.001
Measuring Time	1 to 3 sec
Refractive Index	
Range	1.3330-1.4098
Accuracy	±0.0002
Serum Protin Readout	0.0-12.0
Light Source	Ambient
Sample Volume	0.2 ml
Temperature Range	Operation temperature: 0 °C ( 32 F) - 40 °C ( 104 F).
Calibration	Automatic
Frequency	
Power	Standard 9V battery
Scope Of Delivery	1x Instruction Manual 1x bottle of distill water 1x battery ( 9 v ) 1x Leatheroid carried bag 1x burette

Address. S C O - T e c h  
Auf der Heide 15  
D-37351 Dingelstädt  
Federal Republic of Germany

Phon: + 49 (0) 36075 / 43439-0  
Fax: + 49 (0) 36075 / 43439-9  
E-mail: [sco@sco-tech.com](mailto:sco@sco-tech.com)  
Internet: [www.sco-tech.com](http://www.sco-tech.com)

Contact